

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity modules mentioned in Article 13 of Directive 2014/31/EU, after  
having established that the measuring instrument meets the applicable  
requirements of Directive 2014/31/EU, to:

Manufacturer Changzhou Sunmi Scale Technology Co.,Ltd.  
Xinzha Technology Industrial Park  
Buidling 19&21, 213000  
Changzhou, Jiangu  
China

Measuring instrument **A Non-automatic weighing instrument**  
Type : S2

Further properties are described in the annexes:

- Description T11689 revision 0;
- Documentation folder T11689-1.

Valid until 27 September 2029

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
27 September 2019

  
C. Oosterman  
Head Certification Board

**NMI Certin B.V.**  
Thijssseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision  
that no liability is accepted and that the  
manufacturer shall indemnify third-party  
liability.

The designation of NMI Certin B.V. as Notified  
Body can be verified at  
[http://ec.europa.eu/growth/tools-  
databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete  
document only is permitted.

## 1 General information about the non-automatic weighing instrument

All properties of the non-automatic weighing instrument, whether mentioned or not, shall not be in conflict with the legislation.

### 1.1 Essential parts

The electronics;  
 The mechanical assembly with load cell.

See block diagram;

Number	Pages	Description	Remarks
11689/0-01	1	Block diagram	-

EMI protection measures:


- A/D board shielded with metal cover;
- Ferrite on cable between main board and displays;
- Ferrite on cable between A/D board and load cell;
- Ferrite on power adaptor cable.

### 1.2 Essential characteristics

Accuracy class	$\textcircled{\text{III}}$ or $\textcircled{\text{III}}$
Maximum capacity	$6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval	$e \geq 2 \text{ g}$
Weighing range(s)	Single interval Multi-interval
Maximum number of scale intervals	$n \leq 3000$
Maximum number of weighing ranges	2
Tare	$T \leq -50\% + e$ for instruments with one weighing range $T \leq -\text{Max}_1 + e_1$ for multi-interval instruments
Temperature range	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Power supply voltage	100 – 240 V AC 50/60 Hz
Application	Intended to be used for direct sales to the public
Software identification	Version number 10xxx (x = 0... 9)

The software identification is displayed after selecting:

- "POS Steward" on the main screen;
- Pressing tag "Electronic scale";
- The software identification appears at the right bottom corner.

The non-automatic weighing instrument has loadable software. A CRC checksum is used to ensure the data security. The checksum working status is indicated by a sign  on the screen.

The instrument can be used as a self-service scale.

### 1.3 Essential shapes

Number	Pages	Description	Remarks
11689/0-02	1	Exploded view	-

The data plate is secured against removal by sealing or will be destroyed when removed.

The inscriptions Max, Min, e, as required by Directive 2014/31/EU Annex III point 1.4 are presented in the display by software.

Inside the cabinet is an adjustment lock, located on the A/D board.

### 1.4 Conditional parts

The non-automatic weighing instrument may be equipped with peripheral equipment which is used for the applications listed in Article 1(2), (a) to (f) of Directive 2014/31/EU, provided that the peripheral equipment is certified to be connected to a non-automatic weighing instrument by a Notified Body responsible for type examination under Directive 2014/31/EU, or, that the equipment and the use of the equipment complies with the requirements of WELMEC 2.5 Issue 2 clause 2.2.

The non-automatic weighing instrument may be equipped with a Point Of Sale (POS) or an Electronic Cash Register (ECR), provided that the POS and ECR are certified to be connected to a non-automatic weighing instrument by a Notified Body responsible for type examination under Directive 2014/31/EU.

The non-automatic weighing instrument may be equipped with Point Of Sale (POS) software, provided that the POS software is certified to be run on or connected to a non-automatic weighing instrument by a Notified Body responsible for type examination under Directive 2014/31/EU.

The non-automatic weighing instrument may be equipped with Electronic Funds Transfer equipment (EFT/ECU), if the EFT/ECU represents only the price total on the display.

The non-automatic weighing instrument is fitted with a levelling device and a level indicator, unless the instrument is installed in a fixed position. A ring on the level indicator indicates when the maximum tilt is exceeded.

The non-automatic weighing instrument may be equipped with one or more of the following protective interfaces that have not to be secured:

- RS232;
- Ethernet;
- USB host;
- Audio;
- Drawer kick.

Power supply:

- Brand Jiangsu Chenyang Electron Co.Ltd, type CYSE65-240250.

## 1.5 Non-essential parts

The non-automatic weighing instrument may be connected to non-essential devices, for example but not limited to bar code readers, foot switches, second displays and cash drawers, provided that:

- They do not present primary data used for purposes mentioned in Article 1(2), (a) to (f) of Directive 2014/31/EU unless the "Preliminary observation" in Annex I of the Directive is satisfied;
- They do not lead to an instrument having other essential characteristics than those fixed by this certificate.

Other non-essential parts:

- Internal printer;
- Label printer.

## 2 Information about the main constituent parts of the non-automatic weighing instrument

### 2.1 The electronics

#### 2.1.1 Essential parts

Number	Pages	Description	Remarks
11689/0-03	1	A/D board	Including IC list

#### 2.1.2 Essential characteristics

List of legally relevant functions:

- Determination stability of equilibrium;
- Zero indicating;
- Semi-automatic zero-setting;

- Initial zero-setting;
- Zero-tracking;
- Semi-automatic subtractive tare weighing;
- Preset tare;
- Gravity compensation;
- Adjustment / set-up mode via a switch on the A/D board;
- The adjustment mode is secured with a password;
- Acting upon significant faults;
- Price calculation;
- PLU.

If used for direct sales to the public and equipped with a printer the following legally relevant functions may be present:

- Indications other than primary indications;
- Indication of additional information;
- Memory storage;
- Non-weighed articles;
- Totalisation;
- Price labelling.

### 2.1.3 Conditional parts

Number	Pages	Description	Remarks
11689/0-04	1	Mainboard	Including IC list
11689/0-05	2	Power & Interface board	Including IC list

### 2.1.4 Non-essential parts

Display;  
 Keyboard.

## 2.2 The mechanical assembly with load cell

### 2.2.1 Essential parts

Number	Pages	Description	Remarks
11689/0-06	3	Load cell data sheet	Zemic L6D

## 2.2.2 Essential characteristics

$e \geq E_{\max}/10000$ , or  $e_1 \geq E_{\max}/10000$  for multi-interval instrument;  
 Excitation voltage 5-12 V DC.

## 2.2.3 Essential shapes

Number	Pages	Description	Remarks
11689/0-07	2	Load cell assembly	-

## 3 Seals

To secure components that may not be dismantled or adjusted by the user, the non-automatic weighing instrument has to be secured in a suitable manner on the locations indicated in the drawing:

Number	Pages	Description	Remarks
11689/0-08	2	Sealing	-

## 4 Conditions for conformity assessment

The marks, facilities for the marks and the inscriptions on the non-automatic weighing instrument fulfil the requirements of point 1 of Annex III of Directive 2014/31/EU.

The compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in WELMEC 2, 2015 clause 10, at the time of putting into use.

In case the non-automatic weighing instrument is equipped with a Point Of Sale the conformity assessment procedure of WELMEC 2.2 Issue 3 clause 2.2 shall be taken into account.